



STATE OF WASHINGTON
WASHINGTON STATE PATROL
WASHINGTON STATE TOXICOLOGY LABORATORY

2203 Airport Way South, Suite 360 • Seattle, Washington 98134-2027 • (206) 262-6100 • FAX (206) 262-6145

June 13, 2019

To whom this may concern:

The Washington State Patrol Toxicology Laboratory Division received notification on June 12, 2019 that there is a recall on lot number 8187663 of BD Vacutainer tubes (gray-top tubes). The recall notice states that "this lot has been confirmed to have reduced or no additive within the tube". This lot number may have been contained in blood alcohol collection kits purchased directly by law enforcement agencies from TriTech after September 4, 2018.

Please discontinue use of lot number 8187663 immediately. The official recall notice and a list of all agencies that have submitted evidence to date with this lot number is attached. Contact the Toxicology Laboratory directly for a list of specific agency case numbers, if needed (toxlab@wsp.wa.gov).

Sincerely,

A handwritten signature in blue ink, appearing to read "Fiona J. Couper".

Fiona J. Couper, Ph.D.
Washington State Toxicologist
Washington State Patrol

FJC:bp

